



SGS United Kingdom Ltd.
International Electrical Approvals

South Industrial Estate
Bowburn
Co. Durham
DH6 5AD
United Kingdom
Telephone: +44 (0) 191 377 2000
Fax: +44 (0) 191 377 2020

Uni-Trend International Ltd
Rm 901, 9/F, Nanyang Plaza
57 Hung To Rd
Kwun Tong
Kowloon
Hong Kong

Page 1 of 1

Certificate of Test

Electrical Safety
for

Multimeter / Clamp meter

Date of issue: 17th August 2001

SGS Reference: DUR23330/C/AC/01

Details of sample: Model: UT205 / UT 206
Rated: CAT III 600V, 1000A

Date of receipt: 17th July 2001

Date of test: 23rd July 2001 to 17th August 2001

Test required: Test for compliance with EN61010-1: 1993 Safety requirements for electrical equipment for measurement, control and laboratory use, including A2: 1995, and EN61010-2-032: 1995 Particular requirements for hand-held current clamps, and EN61010-2-031 Particular requirements for hand-held probe assemblies.

Conclusion: The sample tested was found to comply with the standard.

Details of test results are in Job File DUR23330 and are available upon request.

This certificate refers only to the sample submitted for test.

Test Engineer

A Clifford

Authorised Signatory:

R Gough

All enquires relating to this certificate must be directed to the Test Engineer